

## **OSMOCOLL® HNL**

Wescor Cat. No. SS-039

High, Normal, and Low Level

### **Osmolality Control References**

for in vitro diagnostic use

#### **Contents**

This package contains six vials of assayed solution (2 high, 2 normal, and 2 low level) for use as controls for osmolality determination by vapor pressure osmometer.

#### **Storage**

Keep refrigerated – Do not freeze. Under refrigeration, the solution will remain stable until the labeled expiration date. The opened product has a maximum stable life of five days, if kept refrigerated and tightly capped.

#### **Assay Protocol**

For osmolality, each lot is assayed in replicate. Data is obtained from a Wescor vapor pressure osmometer calibrated with Wescor osmolality standards. For colloid osmotic pressure, each lot is assayed in replicate. Data is obtained from a Wescor colloid osmometer that is physically pre-calibrated to a certified manometer. The osmometer is fitted with a 30,000 molecular weight cutoff membrane (Wescor cat. no. SS-030).

#### **Warranty**

Wescor warrants this product to perform as described in its labeling and in Wescor's literature, and disclaims any implied warranty of merchantability or fitness for any other purpose. The warranty is limited to Osmocoll product replacement only.

#### **Assayed Values**

See individual vial labels for assayed osmolality and COP ranges.



Wescor, Inc  
459 South Main Street  
Logan, Utah 84321



MT Promedt Consulting GmbH  
Altenhofstrasse 80  
D-66386 St. Ingbert, Germany

PRINT-0095-01A